

Laboratory Testing Report

ISO 8502-9 Equivalence Validation by NACE SP0508-2010

Prepared for:

HedoN Electronic Developments BV
Elektronicaweg 15, 2628 XG Delft
Netherlands

HedoN Electronic Developments BV Soluble Salt Meter (SSM) Model HED7263901 demonstrated equivalence to the ISO 8502-9 method for determining soluble salts during testing by CCC&L, performed March 4 through September 24, 2009 and October 13, 2014 through January 12, 2015. Testing was performed in accordance with NACE SP0508-2010.

Testing Conducted By:

Corrosion Control Consultants and Labs, Inc.
An independent testing facility
January, 2015

Dates of test performance: 3/4/09 – 9/24/09 and 10/13/14 – 1/12/15



Sarah F Olthof
2015.01.22 17:19:07 -05'00'

Location of testing: Corrosion Control Consultants and Labs, Inc.

4403 Donker Court SE, Grand Rapids, MI 49512